

VERIFICATION OF CALIBRATION
of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program JUN 02 2010

Datamaster cdm S/N 130180

Supervisor/Operator Performing the Verification Procedure:
Name AW Anderson ID# 2467 Date 5/29/10
Agency Seldovia Police Dept. Phone # (907) 274-7640
Instrument Location Seldovia PD

Alco S/N X301530 Target Value .078 High Pressure 500 PSI

Alco Test Value Average .076 .076
1st Alco 2nd Alco

Signature AW Anderson CAB
6/7/10


(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June, 2010.
Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



*of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Datamaster cdm S/N 130180 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180 ✓

MAY 29, 2010 ✓

OPERATOR'S NAME: *COB 6/7/10*
ANDERSON, AU
OPERATOR'S NUMBER: 247
SUBJECT'S LAST NAME:
SIXTY DAY VERIFICATION
SUBJECT'S FIRST NAME/MI :
N/A

O.L. #: 1234567890
DEPT/AGENCY: SOU1
CRGE/REPORT: N/A
TEST TYPE: V
ALCO TARGET VALUE: .078 ✓
ALCO S/N: X301530 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.09 in ✓		
ALCO TARGET	.078	19:29
BLANK TEST	.000	19:30
INTERNAL STANDARD	VERIFIED	19:30
ALCO TV 30.09 in	.076 ✓	19:31
BLANK TEST	.000	19:32
SUBJECT SAMPLE <i>COB 6/7/10</i>	.000	19:32
BLANK TEST	.000	19:33
ALCO TV 30.09 in	.076 ✓	19:33
BLANK TEST	.000	19:34

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180 ✓

MAY 29, 2010 ✓
TIME 19:27

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	30.07 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢