

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130181</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Ronny Simmons</u>	ID# <u>5123</u> Date <u>5-17-12</u>
A Agency <u>AST</u>	Phone # <u>320-6166</u>
Instrument Location <u>Nesana</u>	
Alco S/N <u>0751108</u> ^{4X BMB 7/11/12} Target Value <u>.080</u> ^{per Ronny Simmons 07511080A1 which was Lot 4X} High Pressure <u>400</u>	
B Alco Test Values	<u>.080</u> <u>.080</u> 1 st Alco 2 nd Alco
Signature <u>[Signature]</u> <u>RAST</u>	

(OVER)

BMB
7/11/12

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of July 2012. NAK 4/20/12

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas NIKKI ROTH NAK 4/20/12
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130181

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

MAY 17, 2012

OPERATOR'S NAME: SIMMONS, RONNYA
OPERATOR'S NUMBER: 5123
SUBJECT'S LAST NAME: VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI: A

O.L. #: 1
DEPT/AGENCY: NEND
CASE/REPORT: 1
TEST TYPE: V

ALCO TARGET VALUE: .000
ALCO S/N: 0251108

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.81 in
ALCO TARGET .079 20:40
BLANK TEST .000 20:41
INTERNAL STANDARD VERIFIED 20:41
ALCO TV 29.81 in .000 20:41
BLANK TEST .000 20:42
SUBJECT SAMPLE .000 20:42
BLANK TEST .000 20:43
ALCO TV 29.81 in .000 20:44
BLANK TEST .000 20:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

MAY 17, 2012
TIME 20:37

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 43c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~