

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130181</u> ✓
Name	<u>Malik Jones</u>	ID# <u>5830</u> Date <u>11/26/08</u> ✓
A Agency	<u>AST</u>	Phone # <u>451-5100</u>
Instrument Location	<u>Nenana</u> ✓	
B Alco S/N	<u>X301532</u> ✓	Target Value <u>.079</u> ✓ High Pressure <u>500 PSI</u>
Alco Test Value Average	<u>.078</u> ✓	<u>.078</u> ✓
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		 <u>12/10/08</u>
(OVER)		

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Malik Jones, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



  
Jeanne Swartz

Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

12/10/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130181 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130181

NOVEMBER 26, 2008 ✓  
OPERATOR'S NAME: CB  
JONES/MALIK/JAMAL  
OPERATOR'S NUMBER: 5830  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
V  
O.L. #: 0  
DEPT/AGENCY: NEM1  
CASE/REPORT:  
TEST TYPE: V  
ALCO TARGET VALUE: .079 ✓  
ALCO S/N: K301532

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.59 in		
ALCO TARGET	.078	10:15
BLANK TEST	.000	10:16
INTERNAL STANDARD	VERIFIED	10:16
ALCO TV 29.59 in	.078 ✓	10:16
BLANK TEST	.000	10:17
SUBJECT SAMPLE	.000	10:18
BLANK TEST	.000	10:18
ALCO TV 29.60 in	.078 ✓	10:19
BLANK TEST	.000	10:19

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130181

NOVEMBER 26, 2008 ✓  
TIME 10:23

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ\]^\_`abcdefg|hijklmnop  
qrsuvwxyz{|}~\*