

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130181

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

DECEMBER 21, 2010

OPERATOR'S NAME:

JONES/MALIK/JANAL

OPERATOR'S NUMBER: 5830

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

V

O.L. #: 1

DEPT/AGENCY: NEND

CASE REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .074

ALCO S/N: 78895

--- BREATH ANALYSIS ---

.074 ADJUSTED FOR 29.75 in
ALCO TARGET .073 11:17
BLANK TEST .000 11:18
INTERNAL STANDARD VERIFIED 11:18
ALCO TV 29.75 in .076 11:18
BLANK TEST .000 11:19
SUBJECT SAMPLE .000 11:20
BLANK TEST .000 11:21
ALCO TV 29.77 in .076 11:21
BLANK TEST .000 11:22

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

DECEMBER 21, 2010
TIME 11:41

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 43c
BAROMETER: 29.77 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghi jk lmn
opqrstuvwxyz{|}~