


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: <u>DataMaster cdm S/N 130181</u>	
Name <u>MALIK JONES</u>	ID# <u>5830</u> Date <u>3/16/11</u>
Agency <u>AST</u>	Phone # <u>(907)451-5100</u>
Instrument Location <u>NENANA</u>	
Alco S/N <u>78895</u> Target Value <u>.074</u> High Pressure <u>400 PSI</u>	
Alco Test Value Average <u>.076</u> <u>.075</u>	
1 st Alco 2 nd Alco	
Signature 	
(OVER)	


Handwritten initials and date: BMB 4/1/11


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory. *(Do Not write in the area below)*
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of APRIL, 2011 *own*

 (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130181

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

MARCH 16, 2011

OPERATOR'S NAME: *BMB 4/11/11*
JONES/MALIK/JAHAL
OPERATOR'S NUMBER: 5830
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/NI :
V

O.L. #: 0
DEPT/AGENCY: NEND
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .074
ALCO S/N: 78895

— BREATH ANALYSIS —

.074 ADJUSTED FOR 29.47 in
ALCO TARGET .072 19:45
BLANK TEST .000 19:46
INTERNAL STANDARD VERIFIED 19:46
ALCO TV 29.47 in .076 19:46
BLANK TEST .000 19:47
SUBJECT SAMPLE *BMB 4/11/11* .000 19:48
BLANK TEST .000 19:48
ALCO TV 29.47 in .075 19:49
BLANK TEST .000 19:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

MARCH 16, 2011
TIME 19:51

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.47 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghi jk lmnop
qrs tuvwxyz{|}~*