

JAN 20 2010

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N <u>130182</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Jeffrey ELBIE</u>	ID# <u>6623</u> ✓ Date <u>1-15-11</u> ✓
A Agency <u>Anchorage Police Dept</u>	Phone # <u>786-8900</u>
Instrument Location <u>1400 E 4<sup>th</sup></u>	
B Alco S/N <u>K172971</u> ✓ Target Value <u>.084</u> ✓ <del>.085</del> <sup>11/23/11</sup> High Pressure <u>1000</u>	
Alco Test Values	<u>.087</u> ✓ 1 <sup>st</sup> Alco
	<u>.087</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>[Handwritten Signature]</u>	

(OVER)

(Do Not write in the area below)

*[Handwritten initials]*  
*2/8/11*


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*[Handwritten Signature]*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 11<sup>th</sup> day of FEB, 2011 <sup>2011</sup> ~~2010~~ <sup>DW</sup>

*[Handwritten Signature]* (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 230182

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130182

JANUARY 15, 2011

OPERATOR'S NAME:

ELBIE/JEFFREY/L

OPERATOR'S NUMBER: 6623

SUBJECT'S LAST NAME:

E

SUBJECT'S FIRST NAME/MI :

J/L

O.L. #: 0000

DEPT/AGENCY: ANCL

CASE/REPORT: 11-00

TEST TYPE: V

ALCO TARGET VALUE: .084

ALCO S/N: X172971

## --- BREATH ANALYSIS ---

.084 ADJUSTED FOR 30.29 in  
ALCO TARGET .085 00:23  
BLANK TEST .000 00:24  
INTERNAL STANDARD VERIFIED 00:24  
ALCO TV 30.29 in .087 00:24  
BLANK TEST .000 00:25  
SUBJECT SAMPLE .000 00:26  
BLANK TEST .000 00:26  
ALCO TV 30.29 in .087 00:27  
BLANK TEST .000 00:27

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130182

JANUARY 15, 2011

TIME 00:20

## --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 30.29 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

## PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDE  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklm  
pqrstuvwxyz{|}~^