

MAY 17 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>30182</u>
Name		<u>DAN STURKIE</u>	ID# <u>3559</u> Date <u>5-11-10</u>
A	Agency	<u>Anchorage AK / Anchorage POLICE</u>	Phone # <u>786-8900</u>
Instrument Location <u>Eagle River (AAD) SUBSTATION</u>			
Alco S/N		<u>X124611</u>	Target Value <u>.079</u> High Pressure <u>700</u>
Alco Test Values		<u>.079</u>	<u>.079</u>
		1 st Alco	2 nd Alco
Signature		<u>[Signature]</u>	

[Handwritten initials]
5/24/10

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of June, 2010.

[Signature]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

MAY 11, 2010

OPERATOR'S NAME:
STURKIE

OPERATOR'S NUMBER: 3559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: VOC

DEPT/AGENCY: ANCL

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: X124611

*BMB
5/24/10*

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.64 in		
ALCO TARGET	.078	23:36
BLANK TEST	.000	23:37
INTERNAL STANDARD	VERIFIED	23:37
ALCO TV 29.64 in	.079	23:37
BLANK TEST	.000	23:38
SUBJECT SAMPLE	.000	23:38
BLANK TEST	.000	23:39
ALCO TV 29.62 in	.079	23:39
BLANK TEST	.000	23:40

*BMB
5/24/10*

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

MAY 11, 2010
TIME 23:33

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.64 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~∆