


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 16 2010


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130182</u>
Name	<u>C. RITALA</u>	ID# <u>3791</u> Date <u>6/15/10</u>
Agency	<u>A.P.D.</u>	Phone # <u>786-2640</u>
Instrument Location	<u>ANCHORAGE JAIL</u>	
Alco S/N	<u>2D</u>	Target Value <u>.080</u> High Pressure <u>375</u>
Alco Test Values	<u>081</u> 1 st Alco	<u>082</u> 2 nd Alco
Signature		<u>Ritala/1349</u> <u>088</u> <u>6/15/10</u>

(OVER)


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program



Subscribed and sworn before me this 21st day of July, 2010.

(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N


130182 ✓

(CONTINUED FROM FRONT PAGE)


C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



DIAGNOSTIC CHECK



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

JUNE 15, 2010 ✓

OPERATOR'S NAME:

RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

REPORT OF STANDARD OUT OF RANGE

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: V

ALCO TARGET VALUE: .080 ✓

ALCO S/N: 2D ✓

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.22 in ✓
ALCO TARGET .080 19:42
BLANK TEST .000 19:43
INTERNAL STANDARD VERIFIED 19:43
ALCO TV 30.22 in .081 ✓ 19:43
BLANK TEST .000 19:44
SUBJECT SAMPLE .000 19:45
BLANK TEST .000 19:46
ALCO TV 30.22 in .082 ✓ 19:46
BLANK TEST .000 19:47

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182 ✓

JUNE 15, 2010 ✓

TIME 19:41

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.22 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~Δ