

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130182</u>
A	Name	<u>TIM TUCKWOOD</u>	ID# <u>3446</u> Date <u>7-16-08</u>
	Agency	<u>Alaska state troopers</u>	Phone # <u>895-4800</u>
	Instrument Location	<u>delta court</u>	
B	Alco S/N	<u>X301596</u>	Target Value <u>.081</u> High Pressure <u>950 psi</u>
	Alco Test Value Average	<u>.079</u>	<u>.079</u> (New Alco)
		1 st Alco	2 nd Alco
	Signature	<u>Tim Tuckwood</u>	<u>CSB</u> <u>7/25/08</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that;


(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Tim Tuckwood, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

7/28/08
Date

JUL 21 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

JULY 16, 2008

OPERATOR'S NAME:

TUCKWOOD TIM

OPERATOR'S NUMBER: 3446

SUBJECT'S LAST NAME:

VERIF

SUBJECT'S FIRST NAME/MI :

OF CAL

O.L. #: 12345

DEPT/AGENCY: DJTB

CASE/REPORT: 09876

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: X301596

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 28.97 in
ALCO TARGET .078 09:43
BLANK TEST .000 09:44
INTERNAL STANDARD VERIFIED 09:44
ALCO TV 28.97 in .079 09:44
BLANK TEST .000 09:45
SUBJECT SAMPLE .000 09:46
BLANK TEST .000 09:46
ALCO TV 28.97 in .079 09:47
BLANK TEST .000 09:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

JULY 16, 2008

TIME 09:48

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 28.97 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~