VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 27 2009

DataMaster cdm S/N 130182

Supervisor/Operator Performing the Verification Procedure:

A

Name JSwartz ID# 4314 Date 7/24/09

Agency SCDL Phone # 269-5592

Instrument Location SCDL

B

Alco S/N SCOTT08 Target Value .080 High Pressure 700 psi

Alco Test Values .081 (1st Alco) .083 (2nd Alco)

Signature [signature] 7/30/09

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[signature]

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of Aug, 2009.

[signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182 ✓

JULY 24, 2009 ✓

CDB 7/30/09

OPERATOR'S NAME:
SWARTZ/J
OPERATOR'S NUMBER: 4314
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: AAA1
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .080 ✓
ALCO S/N: SCOTTOS ✓

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.99 in

ALCO TARGET	.080	17:33
BLANK TEST	.000	17:33
INTERNAL STANDARD	VERIFIED	17:33
ALCO TV 29.99 in	.081 ✓	17:34
BLANK TEST	.000	17:34
SUBJECT SAMPLE	.000	17:35
BLANK TEST	.000	17:36
ALCO TV 29.99 in	.083 ✓	17:36
BLANK TEST	.000	17:37

CDB 7/30/09

JJ 7/24/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

JULY 24, 2009 ✓
TIME 17:31

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.99 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
[illegible]