

# VERIFICATION OF CALIBRATION REPORT

JUL 29 2009

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130182</u> ✓	
Name <u>JAMIES MICHELS</u>	ID# <u>6502</u> Date <u>7/25/09</u> ✓
A Agency <u>USBPD</u>	Phone # <u>852-6111</u>
Instrument Location <u>BARROW - USBPD - SQUAD ROOM</u>	
Alco S/N <u>X172971</u> ✓ Target Value <u>.081</u> ✓ High Pressure <u>200</u> <u>PSI</u>	
B Alco Test Values <u>.087</u> ✓ <u>.087</u> ✓	
1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>James Mich</u> ✓	<u>COB</u> <u>8/27/09</u>
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 17<sup>th</sup> day of Sept, 2009.

Carolyn M. Notland (Notary Seal Stamp)

Carolyn M. Notland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130182

JULY 25, 2009 ✓ *CSB*

OPERATOR'S NAME:  
MICHELS/JAMES  
OPERATOR'S NUMBER: 6502  
SUBJECT'S LAST NAME:  
UOC  
SUBJECT'S FIRST NAME/MI :  
UOC  
O.L. #: UOC  
DEPT/AGENCY: BRW1  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: .081 ✓  
ALCO S/N: X172971 ✓

### — BREATH ANALYSIS —

.081 ADJUSTED FOR 30.25 in ✓  
ALCO TARGET .081 21:47  
BLANK TEST .000 21:49  
INTERNAL STANDARD VERIFIED 21:49  
ALCO TV 30.25 in .087 ✓ 21:49  
BLANK TEST .000 21:50  
SUBJECT SAMPLE *CSB* .000 21:50  
BLANK TEST .000 21:51  
ALCO TV 30.25 in .087 ✓ 21:51  
BLANK TEST .000 21:52

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130182

JULY 25, 2009 ✓  
TIME 21:41

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 40c  
BAROMETER: 30.25 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST :

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z I \ ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { } ~ |