VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182

Supervisor/Operator Performing the Verification Procedure:

A

Name Tim Tuckwood ID# 3446 Date 9-1-08

Agency Alaska State Troopers Phone # 895-4800

Instrument Location Delta Court

B

Alco S/N X301596 Target Value .081 High Pressure 900 psi

Alco Test Value Average .078 (1st Alco) .079 (2nd Alco)

Signature Tim Tuckwood

9/11/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Tim Tuckwood , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

9/11/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

SEP 0 8 2008

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

SEPTEMBER 01, 2008 ✓ 9/11/08

OPERATOR'S NAME:
TUCKWOOD TIM
OPERATOR'S NUMBER: 3446
SUBJECT'S LAST NAME:
VERIF
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L. #: 12345
DEPT/AGENCY: DJT0
CASE/REPORT: 09876
TEST TYPE: V
ALCO TARGET VALUE: .081 ✓
ALCO S/N: X301596 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 28.76 in

ALCO TARGET	.077	08:21
BLANK TEST	.000	08:22
INTERNAL STANDARD	VERIFIED	08:22
ALCO TV 28.76 in	.078 ✓	08:22
BLANK TEST	.000	08:23
SUBJECT SAMPLE	.000 9/11/08	08:24
BLANK TEST	.000	08:24
ALCO TV 28.76 in	.079 ✓	08:25
BLANK TEST	.000	08:26

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

SEPTEMBER 01, 2008 ✓
TIME 08:19

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	28.76 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~