

SEP 14 2010

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130182</u> ✓	
Name	<u>Troy Clark</u>	ID#	<u>4934</u> ✓
		Date	<u>09/07/2010</u> ✓
A	Agency <u>Anchorage Police Department</u>	Phone #	<u>786-8900</u>
	Instrument Location <u>Anchorage Jail</u>		
B	Alco S/N <u>2M</u> ✓	Target Value <u>.080</u> ✓	High Pressure <u>1000</u>
	Alco Test Values	<u>.081</u> ✓	<u>.082</u> ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature <u>Troy Clark</u>		

(OVER)

(Do Not write in the area below)

*BMB*  
*9/17/10*

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program



Subscribed and sworn before me this 9th day of NOV, 2010.

*Carolyn M. Noland-DW* (Notary Seal Stamp)  
 Carolyn M. Noland-DW DANIEL L. NICHOLAS  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130182

SEPTEMBER 06, 2010

OPERATOR'S NAME: CLARK/TROY  
OPERATOR'S NUMBER: 4934  
SUBJECT'S LAST NAME: CLARK  
SUBJECT'S FIRST NAME/MI : TROY  
O.L. #: 6633880  
DEPT/AGENCY: ANCI  
CASE/REPORT: VOC  
TEST TYPE: V  
ALCO TARGET VALUE: .080  
ALCO S/N: 2M

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.29 in  
ALCO TARGET .080 02:37  
BLANK TEST .000 02:38  
INTERNAL STANDARD VERIFIED 02:38  
ALCO TV 30.29 in .081 02:38  
BLANK TEST .000 02:39  
SUBJECT SAMPLE .000 02:40  
BLANK TEST .000 02:41  
ALCO TV 30.29 in .082 02:41  
BLANK TEST .000 02:42

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130182

SEPTEMBER 06, 2010  
TIME 02:33

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.29 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~  
^