VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 183

Supervisor/Operator Performing the Verification Procedure:

A

Name HUTT, William T. ID# 3359 Date 01/05/09

Agency Homer Police Dept./Jail Phone # 235-3150

Instrument Location 4060 Heath St. Homer, AK 99603 235-3151 (fax)

B

Alco S/N 75260 Target Value .079 High Pressure 400 psi

Alco Test Value Average .083 (1st Alco) .082 (2nd Alco)

Signature William T. Hutt CDB 1/07/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Hutt , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

2/3/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

JANUARY 05, 2009 CDB 1/7/09

OPERATOR'S NAME:
HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
ETOH
O.L. #: 01234567
DEPT/AGENCY: HOM1
CASE/REPORT: 09-0000
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: 75260

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.59 in

ALCO TARGET	.078	18:43
BLANK TEST	.000	18:44
INTERNAL STANDARD	VERIFIED	18:44
ALCO TU 29.57 in	.083	18:44
BLANK TEST	.000	18:45
SUBJECT SAMPLE	.000	18:46
BLANK TEST	.000	18:47
ALCO TU 29.57 in	.082	18:47
BLANK TEST	.000	18:48

WTH 01/05/09 CDB 1/7/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

JANUARY 05, 2009
TIME 18:39 WTH

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.57 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~