

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAR 09 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130183</u>
Name	<u>HUTT, William T.</u>	ID#	<u>3359</u> Date <u>03/05/09</u>
A Agency	<u>Homer Police Dept. / Jail</u>	Phone #	<u>235-3150</u>
Instrument Location	<u>4060 Heath St. Homer, AK</u>	fax #	<u>235-3151</u>
Alco S/N	<u>75260</u>	Target Value	<u>.079</u> High Pressure <u>300 psi</u>
B Alco Test Value Average	<u>.083</u>	<u>.083</u>	
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature	<u>William T. Hutt</u>		<u>CSB</u> <u>3/24/09</u>

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Hutt, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



J Swartz 3/24/09  
Jeanne Swartz Date  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130183

MARCH 05, 2009 ✓

OPERATOR'S NAME: HUTT, WILLIAM T  
OPERATOR'S NUMBER: 3359

*COB 3/24/09*

SUBJECT'S LAST NAME: VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI: ETON

O.L. #: 01234567  
DEPT/AGENCY: HOM1  
CASE REPORT: 09-0000  
TEST TYPE: V  
ALCO TARGET VALUE: .079 ✓  
ALCO S/N: 75260 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.86 in		
ALCO TARGET	.078	23:16
BLANK TEST	.000	23:17
INTERNAL STANDARD	VERIFIED	23:17
ALCO TV 29.86 in	.083 ✓	23:17
BLANK TEST	.000	23:18
SUBJECT SAMPLE	.000	23:20
BLANK TEST	.000	23:21
ALCO TV 29.88 in	.083 ✓	23:22
BLANK TEST	.000	23:22

*COB 3/24/09*  
*WJH*  
*03/05/09*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130183

MARCH 05, 2009 ✓  
TIME 23:10

*WJH*

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/26/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.86 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~