

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183 ✓

Supervisor/Operator Performing the Verification Procedure:

Name HUTT, William T. ID# 3359 ✓ Date 08/30/10 ✓

A Agency Homer Police Dept./Jail Phone # 235-3150

Instrument Location 4060 Heath St. Homer, AK 99603 235-3151 (fax)

B Alco S/N A090740 ✓ Target Value 0.084 ✓ High Pressure 450 psi

Alco Test Values .084 ✓ .084 ✓
 1st Alco 2nd Alco

Signature William T. Hutt

(OVER)

BMB
9/3/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 5th day of Oct, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

AUGUST 30, 2010

OPERATOR'S NAME:
HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
ETOH

O.L. #: 01234567
DEPT/AGENCY: HOM1
CASE/REPORT: 10-0000
TEST TYPE: U
ALCO TARGET VALUE: 0.084
ALCO S/N: A090740

— BREATH ANALYSIS —

.084 ADJUSTED FOR 29.68 in
ALCO TARGET .083 20:50
BLANK TEST .000 20:51
INTERNAL STANDARD VERIFIED 20:51
ALCO TV 29.68 in .084 20:51
BLANK TEST .000 20:52
SUBJECT SAMPLE WITH 20:53
BLANK TEST .000 20:54
ALCO TV 29.68 in .084 20:55
BLANK TEST .000 20:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

AUGUST 30, 2010
TIME 20:44

WTHOR

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER #: 29.68 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢