

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 04 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130183</u>
Name	<u>HUTT, William T.</u>	ID#	<u>3359</u> Date <u>09/03/09</u>
A Agency	<u>Homer Police Dept./Jail</u>	Phone #	<u>235-3150</u>
Instrument Location <u>406A Heath St. Homer, AK 99603</u>			
Alco S/N		Target Value	High Pressure
<u>X301599</u>		<u>.086</u>	<u>700 psi</u>
Alco Test Values		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
<u>.087</u>		<u>.086</u>	
Signature <u>William T. Hutt</u>			

(OVER)

BMB  
9/15/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 18<sup>th</sup> day of Sept, 2009.

Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130183

SEPTEMBER 03, 2009

OPERATOR'S NAME:  
HUTT, WILLIAM T  
OPERATOR'S NUMBER: 3359  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
ETOH  
O.L. #: 01234567  
DEPT/AGENCY: HOM1  
CASE/REPORT: 09-0000  
TEST TYPE: U  
ALCO TARGET VALUE: .006  
ALCO S/N: X301599

### — BREATH ANALYSIS —

.006 ADJUSTED FOR 29.60 in ✓  
ALCO TARGET .005 00:49  
BLANK TEST .000 00:50  
INTERNAL STANDARD VERIFIED 00:50  
ALCO TU 29.60 in .007 00:51  
BLANK TEST .000 00:51  
SUBJECT SAMPLE .000 00:53  
BLANK TEST .000 00:54  
ALCO TU 29.60 in .006 00:54  
BLANK TEST .000 00:55

*Handwritten:* 9/15/09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130183

SEPTEMBER 03, 2009  
TIME 00:46

*Handwritten:* WTH

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.60 in ✓  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLmnopqrstuvwxyz{|}~`\_.'abcde/ghijklmno  
pqrstuvwxyz{|}~`\_.