

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130183</u> ✓
Name		<u>HUTT, William T.</u>	ID# <u>3359</u> Date <u>09/05/08</u> ✓
A	Agency	<u>Homer Police Dept./Jail</u>	Phone # <u>235-3150</u>
Instrument Location		<u>4060 Heath St. Homer, AK 9960</u> ✓	
Alco S/N		<u>75260</u> ✓	Target Value <u>.079</u> ✓ High Pressure <u>750 psi</u>
B	Alco Test Value Average	<u>.083</u> ✓ 1 st Alco	<u>.082</u> ✓ 2 nd Alco
Signature		<u>William T. Hutt</u> ✓	<u>005</u> <u>9/11/08</u>

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Hutt, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

9/11/08

Date

BT9 03/00

SEP 08 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

PAGE

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

SEPTEMBER 05, 2008

OPERATOR'S NAME:
HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
ETOH

O.L. #: 01234567
DEPT/AGENCY: HOM1
CASE/REPORT: 08-0000
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: 75260

BREATH ANALYSIS

.079 ADJUSTED FOR 30.04 in
ALCO TARGET .079 00:32
BLANK TEST .000 00:33
INTERNAL STANDARD VERIFIED 00:33
ALCO TV 30.03 in .083 00:33
BLANK TEST .000 00:34
SUBJECT SAMPLE .000 00:35
BLANK TEST .000 00:36
ALCO TV 30.04 in .082 00:37
BLANK TEST .000 00:37

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

SEPTEMBER 05, 2008
TIME 00:29

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.03 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~