



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130183

DECEMBER 29, 2010

OPERATOR'S NAME: HUTT, WILLIAM T  
OPERATOR'S NUMBER: 3359  
SUBJECT'S LAST NAME: VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI : ETON  
O.L. #: 01234567  
DEPT/AGENCY: ADM1  
CASE/REPORT: 10-0000  
TEST TYPE: U  
ALCO TARGET VALUE: 0.076  
ALCO S/N: X124630

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.36 in  
ALCO TARGET .074 17:49  
BLANK TEST .000 17:50  
INTERNAL STANDARD VERIFIED 17:50  
ALCO TV 29.36 in .077 17:51  
BLANK TEST .000 17:51  
SUBJECT SAMPLE .000 17:53  
BLANK TEST .000 17:54  
ALCO TV 29.36 in .078 17:54  
BLANK TEST .000 17:55

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130183

DECEMBER 29, 2010  
TIME 17:41

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.36 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKL NOPQRSTUVWXYZ[\]^\_`'abcde fghijklmno  
pqrstuvwxy z{|}~