

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

MARCH 09, 2011

OPERATOR'S NAME:
HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
FINN
O.L. #: 01234567
DEPT/AGENCY: HOM1
CASE/REPORT: 11-0000
TEST TYPE: V
ALCO TARGET VALUE: 0.076
ALCO S/N: X124630

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.65 in
ALCO TARGET .075 20:39
BLANK TEST .000 20:40
INTERNAL STANDARD VERIFIED 20:40
ALCO TV 29.65 in .078 20:40
BLANK TEST .000 20:41
SUBJECT SAMPLE .000 20:42
BLANK TEST .000 20:43
ALCO TV 29.65 in .078 20:43
BLANK TEST .000 20:44

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

MARCH 09, 2011
TIME 20:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.65 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!*"#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL*NOPQRSTUWXYZ!*^_`'abcde fghi jklmno
pqrstuvwxyz{|}~@

WTHA

BOMB
4/1/11

WTHA

WTHA
3/31/11

BOMB
4/1/11