

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130181</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>JOHN HAMILTON</u>	ID# <u>3192</u> Date <u>5/20/12</u>	
A Agency <u>PETERSBURG POLICE DEPARTMENT</u> Phone # <u>772-3838</u>		
Instrument Location <u>PETERSBURG</u>		
B Alco S/N <u>4K</u> Target Value <u>.080</u> High Pressure <u>800 PSI</u>		
Alco Test Values	<u>.078</u> 1st Alco	<u>.078</u> 2nd Alco
Signature		


(OVER)

(Do Not write in the area below)

*Handwritten:* BMB 6/14/12


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 14 day of June, 2012. *NAK 6/14/12*

  
(Notary Seal Stamp)  
Daniel L. Nicholas *Nikki Roth 6/14/12 NAK*  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130184 *BMB 6/14/12*

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130184

MAY 20, 2012

OPERATOR'S NAME: *BMB 6/14/12*  
HAMILTON JOHN A  
OPERATOR'S NUMBER: 3192  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/NI :  
A  
O.L. #: A  
DEPT/AGENCY: PSG1  
CASE/REPORT: A  
TEST TYPE: U  
ALCO TARGET VALUE: .000  
ALCO S/N: 4K

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.01 in  
ALCO TARGET .000 10:07  
BLANK TEST .000 10:08  
INTERNAL STANDARD VERIFIED 10:09  
ALCO TV 30.01 in .078 10:09  
BLANK TEST .000 10:10  
SUBJECT SAMPLE *BMB 6/14/12* .000 10:10  
BLANK TEST .000 10:11  
ALCO TV 30.01 in .078 10:12  
BLANK TEST .000 10:12

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130184

MAY 20, 2012  
TIME 10:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 40c  
BAROMETER: 30.01 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!""#%&'()\*+,-./0123456789:;<=>?@ABCDEFG  
HIJKL MNOPQRSTUVWXYZ[\]^\_`abcde fghi jklmno  
pqrs tuvwxyz{|}~■