VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184

Supervisor/Operator Performing the Verification Procedure:

Name JOHN HAMILTON ID# 3192 Date 2/5/09

A

Agency PETERSBURG POLICE DEPARTMENT Phone # 772-3838

Instrument Location PETERSBURG

B

Alco S/N X301600 Target Value .085 High Pressure 800 PSI

Alco Test Values .086 (1st Alco) .086 (2nd Alco)

Signature [signature] CAB 2/18/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Hamilton, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 2/18/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

FEB 09 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130184

FEBRUARY 05, 2009 - CCB 2/18/09

OPERATOR'S NAME:
HAMILTON JOHN A
OPERATOR'S NUMBER: 3192
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
P
O.L. #: A
DEPT/AGENCY: PSG1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .085
ALCO S/N: X301600

--- BREATH ANALYSIS ---

.085 ADJUSTED FOR 30.12 in

ALCO TARGET	.085	23:15
BLANK TEST	.000	23:16
INTERNAL STANDARD	VERIFIED	23:16
ALCO TV 30.12 in	.086	23:17
BLANK TEST	.000	23:18
SUBJECT SAMPLE CCB 2/18/09	.000	23:19
BLANK TEST	.000	23:19
ALCO TV 30.14 in	.086	23:20
BLANK TEST	.000	23:21

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130184

FEBRUARY 05, 2009
TIME 23:22

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.14 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~