

VERIFICATION OF CALIBRATION REPORT

of DataMaster edm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster edm S/N	<u>130184</u> ✓
Name	<u>JOHN HAMILTON</u> ✓	ID#	<u>3192</u> Date <u>12/5/08</u> ✓
A	Agency	<u>PETERSBURG POLICE DEPARTMENT</u>	Phone # <u>772-3838</u>
	Instrument Location	<u>PETERSBURG PD</u> ✓	
Alco S/N		<u>X 301600</u> ✓	Target Value <u>.085</u> ✓ High Pressure <u>900 PSI</u>
B		Alco Test Values	
		<u>.086</u> ✓	<u>.087</u> ✓
		1 st Alco	2 nd Alco
Signature	 ✓		<u>COB</u> <u>12/18/08</u>


(OVER)


(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Hamilton, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

12/20/08

BT9 03/00

DEC 09 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130184

DECEMBER 05, 2008

OPERATOR'S NAME:
HAMILTON JOHN A
OPERATOR'S NUMBER: 3192
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT AGENCY: P961
CRIME/REPORT: A
TEST TYPE: U
ALCO TARGET VALUE: .005 ✓
ALCO S/N: X301600

--- BREATH ANALYSIS ---

.005 ADJUSTED FOR 29.94 in
ALCO TARGET .005 05:32
BLANK TEST .000 05:33
INTERNAL STANDARD VERIFIED 05:33
ALCO TO 29.94 in .006 ✓ 05:34
BLANK TEST .000 05:34
SUBJECT SAMPLE .000 05:35
BLANK TEST .000 05:36
ALCO TO 29.94 in .007 ✓ 05:36
BLANK TEST .000 05:37

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130184

DECEMBER 05, 2008 ✓
TIME 05:39

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHANGER: 49c
BREATH TUBE: 42c
BAROMETER: 29.94 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! "# \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~