

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 30 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130187 ✓
Name JOHN HAMILTON ✓ ID# 3192 ✓ Date 4/25/10 ✓
A Agency PETERSBURG POLICE DEPARTMENT Phone # 772-3838
Instrument Location PETERSBURG
B Alco S/N 78820 ✓ Target Value .083 High Pressure 850 PSI
Alco Test Values .084 .085
1st Alco 2nd Alco
Signature [Signature]
(OVER)

BMB
6/23/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program



Subscribed and sworn before me this 2nd day of July, 2010.
[Signature]
(Notary Seal Stamp)
Carolyn M. Noland

Notary Public, State of Alaska
Commission Expires with Office



Target value not programmed. Breath
Personnel needed to authenticate voc
for court. BMB 6/23/10

B 19 06/03

