

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130185
Name	Brandi Barnett	ID#	7038
		Date	06/04/2010 <sup>6/3/10 BMB</sup>
A	Agency	SCDL	Phone # 269-5695
Instrument Location		SCDL	
Alco S/N		Scott08	Target Value 0.080
			High Pressure 450 psi
B	Alco Test Values		0.079
		1 <sup>st</sup> Alco	0.078
		2 <sup>nd</sup> Alco	
Signature		<i>Brandi M. Barnett</i>	
		CMB 6/21/10	

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2<sup>nd</sup> day of July, 2010.

*Carolyn M. Noland*  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

JUNE 03, 2010 *CBS 6/2/10*

OPERATOR'S NAME:  
BARNETT/B

OPERATOR'S NUMBER: 7038

SUBJECT'S LAST NAME:  
VOC

SUBJECT'S FIRST NAME/MI :  
VOC

O.L. #: VOC

DEPT/AGENCY: AAA1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: 0.090

ALCO S/N: SCOTT08

--- BREATH ANALYSIS ---

.090 ADJUSTED FOR 29.59 in		
ALCO TARGET	.079	08:51
BLANK TEST	.000	08:52
INTERNAL STANDARD	VERIFIED	08:52
ALCO TV 29.59 in	.079	08:52
BLANK TEST	.000	08:53
SUBJECT SAMPLE	.000	08:54
BLANK TEST	.000	08:54
ALCO TV 29.59 in	.078	08:55
BLANK TEST	.000	08:55

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

JUNE 03, 2010  
TIME 08:49

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 41c

BAROMETER: 29.59 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z ( ) + \*