

JUL 01 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130185</u> ✓
Name <u>Todd W. Bruno</u>	ID# <u>7672</u> ✓	Date <u>06/25/2010</u> ^{6/26/2010} _{7/01/2010} ✓
A Agency <u>Glacier Bay National Park</u>		Phone # <u>(907) 697-2687</u>
Instrument Location <u>Bartlett Cove Ranger Station</u>		
Alco S/N <u>2H</u> ✓		Target Value <u>0.080</u> ✓
		High Pressure <u>1000 psi</u>
Alco Test Values <u>0.081</u> ✓		<u>0.081</u> ✓
1 st Alco		2 nd Alco
Signature <u>Todd Bruno</u>		<u>003</u> <u>7/15/10</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 26th day of July, 2010.

Carolyn M. Noland
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130185
 JUNE 26, 2010 *COB #115110*
 OPERATOR'S NAME:
 BRUNO/TODD/W
 OPERATOR'S NUMBER: 7672
 SUBJECT'S LAST NAME:
 TEST
 SUBJECT'S FIRST NAME/MI :
 TEST
 O.L. #: 0000
 DEPT/AGENCY: GL3A
 CASE/REPORT: 0000
 TEST TYPE: V
 ALCO TARGET VALUE: 0.252
 ALCO S/N: 2H

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.20 in
 ALCO TARGET .000 11:15
 BLANK TEST .000 11:16
 INTERNAL STANDARD VERIFIED 11:16
 ALCO TV 30.22 in .051 11:16
 BLANK TEST .000 11:17
 SUBJECT SAMPLE *COB #115110* .000 11:17
 BLANK TEST .000 11:18
 ALCO TV 30.22 in .051 11:18
 BLANK TEST .000 11:19

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130185

JUNE 26, 2010
 TIME 10:21

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
 PROGRAM: OKAY
 SOFTWARE DATE: 01/28/09
 HEATERS
 SAMPLE CHAMBER: 43c
 BREATH TUBE: 43c
 BAROMETER: 30.16 in
 FLOW DETECTOR: OKAY
 PUMP
 HIGH SPEED: OKAY
 DETECTOR: OKAY
 FILTERS: OKAY
 QUARTZ STANDARD: OKAY
 CALIBRATION: OKAY

PRINTER TEST

!##%&'()*+,-./0123456789:;<=>?@ABCDEF
 HIJKL MNOPQRSTUVWXYZ [\] ^ _ ` abcdefghijklmnop
 qrstuvwxyz{|}~@