SEP 1 6 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

Supervisor/Operator Performing the Verification Procedure:

Name TODD BRUNO ID# 7672 Date 9/11/10

A

Agency NPS GLACIER BAY NP&P Phone # 697-2687

Instrument Location BARTLETT COVE RANGER STATION

B

Alco S/N 2K Target Value 0.080 High Pressure 1000 psi

Alco Test Values 0.079 (1st Alco) 0.080 (2nd Alco)

Signature Todd Bruno

(OVER)

BMB 9/23/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of Nov, 2010.

(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

DANIEL L. NICHOLAS
Commission # 116780
NOTARY PUBLIC
My Commission Expires With Office
STATE OF ALASKA

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

SEPTEMBER 11, 2009

BMB 9/23/10

OPERATOR'S NAME:
BRUNO/TODD/W
OPERATOR'S NUMBER: 7672
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/MI :
TEST
O.L. #: 0000
DEPT/AGENCY: NPS/GLBA
CASE/REPORT: 0002
TEST TYPE: V
BLOOD TARGET VALUE: 0.352
BLOOD S/N: 2K

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.49 in

BLOOD TARGET	.081	11:13
BLANK TEST	.000	11:14
INTERNAL STANDARD	VERIFIED	11:14
BLOOD TV 30.48 in	.079	11:15
BLANK TEST	.000	11:15
SUBJECT SAMPLE	[illegible]	[illegible]
BLANK TEST	.000	[illegible]
BLOOD TV 30.48 in	.[illegible]	[illegible]
BLANK TEST	.000	11:18

BMB 9/23/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

SEPTEMBER 11, 2009
TIME 11:09

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	49c
BAROMETER:	30.49 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
[illegible]
pqrstuvwxyz{|}~