

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130185 ^{BMB} 2/8/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

JANUARY 08, 2011

OPERATOR'S NAME:
UMRS/CHRISTOPHER/J
OPERATOR'S NUMBER: 3553
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: VOC
DEPT/AGENCY: NPS/GLB-
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .055
ALCO S/N: 2K

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.64 in
ALCO TARGET .052 09:59
BLANK TEST .000 10:00
INTERNAL STANDARD VERIFIED 10:00
ALCO TO 30.64 in .079 10:00
BLANK TEST .000 10:01
SUBJECT SAMPLE .000 10:01
BLANK TEST .000 10:02
ALCO TO 30.64 in .079 10:02
BLANK TEST .000 10:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

JANUARY 08, 2011
TIME 09:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/30/09
HEATERS
SAMPLE CHARGER: 43c
BREATH TUBE: 41c
BAROMETER: 30.64 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./:;<=>?@A B C D E F G H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o p q r s t u v w x y z { } ~