

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 09 2009

DataMaster cdm S/N <u>130186</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Bill J Ferris</u>	ID# <u>5004</u> Date <u>3/1/09</u> ✓
A Agency <u>Togiak Police Department</u>	Phone # <u>493-5212</u>
Instrument Location <u>100 Main Street Togiak</u> ✓	
Alco S/N <u>57023</u> ✓ Target Value <u>081</u> ✓ High Pressure <u>1200 psi</u>	
B Alco Test Value Average <u>078</u> ✓ <u>078</u> ✓	
1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>William J Ferris</u> ✓	
<u>CSB</u> <u>3/24/09</u>	

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bill J Ferris, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

3/24/09  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130186 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130186

MARCH 01, 2009

OPERATOR'S NAME: *Bill Ferris*  
FERRIS BILL J  
OPERATOR'S NUMBER: 5004  
SUBJECT'S LAST NAME:

UOC  
SUBJECT'S FIRST NAME/MI :  
XXXX

O.L. #: 123-456  
DEPT/AGENCY: T061  
CASE/REPORT: XX-XXX  
TEST TYPE: U  
ALCO TARGET VALUE: 081 ✓  
ALCO S/N: 57023 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.42 in		
ALCO TARGET	.082	20:48
BLANK TEST	.000	20:48
INTERNAL STANDARD	VERIFIED	20:49
ALCO TV 30.42 in	.078 ✓	20:49
BLANK TEST	.000	20:50
SUBJECT SAMPLE	.000 <i>Bill Ferris</i>	20:50
BLANK TEST	.000	20:51
ALCO TV 30.42 in	.078 ✓	20:51
BLANK TEST	.000	20:52

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130186

MARCH 01, 2009 ✓  
TIME 28:44

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.42 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghi jklmno  
pqrstuvwxyz{|}~■