

AUG 27 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130186 ✓

Name Byron D. Redburn ID# 2087 Date 8-17-2010 ✓

A Agency Nome PD Phone # 443-5262

Instrument Location Nome

---

B Alco S/N X301548 ✓ Target Value -0.79 ✓ High Pressure 300

Alco Test Values .080 ✓  
~~0.78~~ 0.78 <sup>0.78</sup> <sub>8/27/10</sub> -0.98 ✓  
 1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Signature]

(OVER)

(Do Not write in the area below)

*BMB*  
*8/20/10*

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program



Subscribed and sworn before me this 23<sup>RD</sup> day of SEPT, 2010.

[Signature] (Notary Seal Stamp)

Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of the Department of Public Safety  
State of Alaska

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130186

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130186

AUGUST 17, 2010

*BMB*  
*8/30/10*

OPERATOR'S NAME:  
REDBURN BYRON D  
OPERATOR'S NUMBER: 2087  
SUBJECT'S LAST NAME:  
NON-DRINKING  
SUBJECT'S FIRST NAME/MI :  
VERIFICATION OF CAL  
O.L. #: A  
DEPT/AGENCY: ONE1  
CASE/REPORT: A  
TEST TYPE: U  
ALCO TARGET VALUE: .079  
ALCO S/N: X301548

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.86 in		
ALCO TARGET	.078	08:49
BLANK TEST	.000	08:50
INTERNAL STANDARD	VERIFIED	08:50
ALCO TU 29.86 in	.000	08:51
BLANK TEST	.000	08:51
SUBJECT SAMPLE	.000	08:53
BLANK TEST	.000	08:54
ALCO TU 29.86 in	.000	08:54
BLANK TEST	.000	08:55

*BMB*  
*8/30/10*

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130186

AUGUST 17, 2010  
TIME 08:45

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.86 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>@ABCDEFGHIJK  
LMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~