

VERIFICATION OF CALIBRATION REPORT
by DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 08 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130186 ✓
Name Byron D. Redburn ID# 2097 Date 10-31-10 ✓
A Agency Nome PD Phone # 443-5262
Instrument Location Nome PD
B Alco S/N X301548 ✓ Target Value .079 ✓ High Pressure 200
Alco Test Values .079 ✓ .079 ✓
1st Alco 2nd Alco
Signature [Signature]
(OVER)

BMB
12/9/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of DEC, 2010.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

BT9 06/03



VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instruments
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130186

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130186

OCTOBER 31, 2010

OPERATOR'S NAME: *Bmb 12/9/10*
FEDBURN BYRON D
OPERATOR'S NUMBER: 2887
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT/AGENCY: OME1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X381548

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.49 in		
ALCO TARGET	.077	12:27
BLANK TEST	.000	12:28
INTERNAL STANDARD	VERIFIED	12:28
ALCO TV 29.49 in	.079	12:29
BLANK TEST	.000	12:30
SUBJECT SAMPLE	.000	12:30
BLANK TEST	.000	12:31
ALCO TV 29.49 in	.079	12:31
BLANK TEST	.000	12:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130186

OCTOBER 31, 2010
TIME 12:25

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/89
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	43c
BAROMETER:	29.49 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~||