

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>DAV PASQUARICUO</u>	DataMaster cdm S/N <u>134187</u> ID# <u>2620</u> Date <u>1-21-12</u>
A Agency <u>DILLINGHAM PD</u>	Phone # <u>842-5354</u>
Instrument Location <u>DILLINGHAM PD</u>	<u>0.080 (NJB) 2/2/2012</u>
B Alco S/N <u>58428</u>	Target Value <u>.080</u> ^{0.077} High Pressure <u>950 PSI</u> _{CRF 1/31/12}
Alco Test Values	<u>.073</u> <u>.073</u> 1st Alco 2nd Alco
Signature <u>[Signature]</u>	
(OVER) <u>CRF 1/31/12</u>	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of Feb, 2012 CRF

[Signature] (Notary Seal Stamp)

~~Daniel L. Nicholas~~ Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130187

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

JANUARY 21, 2012

CRF
1/31/12

OPERATOR'S NAME:
PASQUARIELLO, DAN
OPERATOR'S NUMBER: 2620
SUBJECT'S LAST NAME:
UIC
SUBJECT'S FIRST NAME/MI :
UIC
O.L. #: 1234567890
DEPT/AGENCY: DL61
CASE/REPORT: 1234567890
TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: 58428

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 28.97 in		
ALCO TARGET	.077	18:54
BLANK TEST	.000	18:55
INTERNAL STANDARD	VERIFIED	18:55
ALCO TV 28.97 in	.073	18:55
BLANK TEST	.000	18:56
SUBJECT SAMPLE	.000	18:56
BLANK TEST	.000	18:57
ALCO TV 28.97 in	.073	18:58
BLANK TEST	.000	18:59

CRF
1/31/12

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

JANUARY 21, 2012
TIME 18:59

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	40c
BREATH TUBE:	41c
BAROMETER:	28.97 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrs tuvwxyz{|}~■

