VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187

Supervisor/Operator Performing the Verification Procedure:

A

Name DAN PASQUARIELLO ID# 2620 Date 5-25-12

Agency DILLINGHAM PD Phone # 842-5354

Instrument Location DILLINGHAM PD

B

Alco S/N 58428 Target Value .080 High Pressure 800 PS.

Alco Test Values .080 (1st Alco) .080 (2nd Alco)

Signature

(OVER)

BMB 7/5/12

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of July, ~~2010.~~ 2012 NAR 9/20/12

Nikki Roth (Notary Seal Stamp)
~~Daniel L. Nicholas~~ Nikki Roth NAR 9/20/12
Notary Public, State of Alaska
Commission Expires with Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

MAY 25, 2012 BMB 7/5/12

OPERATOR'S NAME:
PASQUARIELLO, DAN
OPERATOR'S NUMBER: 2620
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 1234567890
DEPT/AGENCY: DLG1
CASE/REPORT: 1234567890
TEST TYPE: U
ALCO TARGET VALUE: .080
ALCO S/N: 58428

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.83 in

ALCO TARGET	.079	12:46
BLANK TEST	.000	12:47
INTERNAL STANDARD	VERIFIED	12:47
ALCO TV 29.83 in	.080	12:48
BLANK TEST	.000	12:48
SUBJECT SAMPLE	.000	12:50
BLANK TEST	.000	12:50
ALCO TV 29.83 in	.080	12:51
BLANK TEST	.000	12:52

BMB 7/5/12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

MAY 25, 2012
TIME 12:52

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.83 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~