

**VERIFICATION OF CALIBRATION REPORT**  
*of DataMaster cdm Breath Test Instrument*  
*State of Alaska*  
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130187</u> ✓
Name <u>Dan Pasquariello</u>	ID# <u>2620</u>	Date <u>5-18-07</u> ✓
A Agency <u>Dillingham PD</u> ✓		Phone # <u>842-5354</u>
Instrument Location <u>Dillingham PD</u>		
B Alco S/N <u>X301550</u> ✓		Target Value <u>.078</u> ✓ High Pressure <u>400 psi</u>
Alco Test Values	<u>.079</u> ✓ 1 <sup>st</sup> Alco	<u>.079</u> ✓ 2 <sup>nd</sup> Alco
Signature <u><i>[Handwritten Signature]</i></u>		<u><i>[Handwritten Signature]</i></u> 6/5/08

Under the Alaska Rules of Evidence, I certify that (OVER)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daniel Pasquariello Jr., whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



*[Handwritten Signature]* 6/9/08  
 Jeanne Swartz Date  
 Scientific Director Designee  
 State of Alaska  
 State Breath Alcohol Testing Program

MAY 27 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130187

MAY 18, 2008

OPERATOR'S NAME:

PASQUARIELLO, DAN

OPERATOR'S NUMBER: 2620

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VERIFICATION OF CAL

O.I.L. #: 1234567890

DEPT/AGENCY: DLG1

CASE/REPORT: 1234567890

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X301550

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.27 in  
ALCO TARGET .078 20:28  
BLANK TEST .000 20:29  
INTERNAL STANDARD VERIFIED 20:29  
ALCO TV 30.27 in .079 20:30  
BLANK TEST .000 20:30  
SUBJECT SAMPLE .000 20:31  
BLANK TEST .000 20:32  
ALCO TV 30.27 in .079 20:32  
BLANK TEST .000 20:33

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130187

MAY 18, 2008

TIME 20:33

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.27 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrs tuvwxyz{|}~