

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska  
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

**APR 26 2010**

DataMaster cdm S/N <u>130187</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Dan Pasquariello</u>	ID# <u>2620</u> Date <u>3-29-10</u>
A Agency <u>Dillingham PD</u>	Phone # <u>842-5354</u>
Instrument Location <u>Dillingham PD</u>	
B Alco S/N <u>78853</u> Target Value <u>.083</u> High Pressure <u>500 psi</u>	
Alco Test Value Average <u>.80</u>	<u>.079</u>
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u><i>[Signature]</i></u>	
(OVER)	

*BMB*  
*4/28/10*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*[Signature]*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 25<sup>th</sup> day of April, 2010.

*[Signature]* (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130187

MARCH 29, 2010 ✓

OPERATOR'S NAME: *BMMB*  
PASQUARIELLO, DAN  
OPERATOR'S NUMBER: 2628  
SUBJECT'S LAST NAME: *4/28/10*  
UOC  
SUBJECT'S FIRST NAME/MI :  
UOC  
O.L. #: 1234567890  
DEPT/AGENCY: DL61  
CASE/REPORT: 1234567890  
TEST TYPE: T  
ALCO TARGET VALUE: .003 ✓  
ALCO S/N: 78853 ✓

### — BREATH ANALYSIS —

.003 ADJUSTED FOR 29.18 in ✓		
ALCO TARGET	.000	18:33
BLANK TEST	.000	18:34
INTERNAL STANDARD	VERIFIED	18:34
ALCO TU 29.20 in	.000	18:34
BLANK TEST	.000 ✓	18:35
SUBJECT SAMPLE <i>BMMB</i>	.000	18:35
BLANK TEST	.000	18:36
ALCO TU 29.20 in	.079	18:37
BLANK TEST	.000	18:37

*4/28/10*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130187 ✓

MARCH 29, 2010 ✓  
TIME 18:38

### — DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	43c
BAROMETER:	29.20 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!\*"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~