


AUG 02 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130187</u> ✓	
Name	<u>Dan Pasquariello</u>	ID#	<u>2620</u> ✓
		Date	<u>7-27-10</u> ✓
A Agency	<u>Dillingham PD</u>	Phone #	<u>842-5354</u>
Instrument Location <u>Dillingham PD</u>			
B Alco S/N	<u>78853</u> ✓	Target Value	<u>.083</u> ✓
		High Pressure	<u>200 psi</u>
Alco Test Values	<u>.083</u> ✓	<u>.082</u> ✓	
	1st Alco	2nd Alco	
Signature			

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 11th day of AUGUST, 2010.

 (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

JULY 27, 2010

OPERATOR'S NAME: *COB 8/9/10*
PASQUARIELLO, DAN
OPERATOR'S NUMBER: 2620
SUBJECT'S LAST NAME:

UOC
SUBJECT'S FIRST NAME/MI :
UOC

D.L. #: 1234567890
DEPT/AGENCY: DL61
CASE/REPORT: 1234567890
TEST TYPE: U
ALCO TARGET VALUE: .083
ALCO S/N: 78853

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.14 in
ALCO TARGET .083 17:19
BLANK TEST .000 17:20
INTERNAL STANDARD VERIFIED 17:20
ALCO TV 30.14 in .083 17:21
BLANK TEST .000 17:21
SUBJECT SAMPLE *COB 8/9/10* .000 17:22
BLANK TEST .000 17:23
ALCO TV 30.14 in .082 17:23
BLANK TEST .000 17:24

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

JULY 27, 2010
TIME 17:24

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 30.14 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! "# \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢