

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130187</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Dan Pasquariello</u>	ID# <u>2620</u> ✓
Date <u>9-17-10</u> ^{OLW} <u>9/28/10</u>	
A Agency <u>Dillingham PD</u>	Phone # <u>842-5354</u>
Instrument Location <u>Dillingham PD</u>	
B Alco S/N <u>70108</u> ✓ Target Value <u>.081</u> ✓ High Pressure <u>1050 psi</u>	
Alco Test Values	<u>.083</u> ✓ 1st Alco
	<u>.084</u> ✓ 2nd Alco
Signature <u></u>	


(OVER)

*BMB
9/28/10*


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

 (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

SEPTEMBER 16, 2010

OPERATOR'S NAME:

PASQUARIELLO, DAN

OPERATOR'S NUMBER: 2620

SUBJECT'S LAST NAME:

UC

SUBJECT'S FIRST NAME/MI :

UC

O.L. #: 1234567890

DEPT/AGENCY: DL61

CASE/REPORT: 1234567890

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: 70100

— BREATH ANALYSIS —

.001 ADJUSTED FOR 30.35 in
ALCO TARGET .002 20:01
BLANK TEST .000 20:02
INTERNAL STANDARD VERIFIED 20:02
ALCO TV 30.35 in .003 20:03
BLANK TEST .000 20:03
SUBJECT SAMPLE .000 20:04
BLANK TEST .000 20:05
ALCO TV 30.35 in .004 20:05
BLANK TEST .000 20:06

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

SEPTEMBER 16, 2010

TIME 20:07

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.35 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLmnopqrstuvwxyz\]^_`abcdefgijklmno
pqrstuvwxyz{|}~