

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130188 ✓

Name V. Eugene Fowler ID# 4386 ✓ Date 8-21-10 ✓

A Agency Alaska State Troopers Phone # 262-4453

Instrument Location Anchor Point AST

B Alco S/N X124635 ✓ Target Value .080 ✓ High Pressure 1025

Alco Test Values .084 ✓ .085 ✓
1st Alco 2nd Alco

Signature V E Fowler

(OVER)

marked on 8-23-10 VE

BMB 9/27/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program



Subscribed and sworn before me this 5th day of Oct, 2010.

Daniel L. Nicholas (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130188

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188

AUGUST 21, 2010

OPERATOR'S NAME:
FOWLER/V/E
OPERATOR'S NUMBER: 4386
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

D.L. #: VOC
DEPT/AGENCY: SLTS
CASE/REPORT: VOC
TEST TYPE: T
ALCO TARGET VALUE: .000
ALCO S/N: X124635

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.04 in
ALCO TARGET .000 22:07
BLANK TEST .000 22:08
INTERNAL STANDARD VERIFIED 22:08
ALCO TV 30.04 in .004 22:08
BLANK TEST .000 22:09
SUBJECT SAMPLE .000 22:10
BLANK TEST .000 22:10
ALCO TV 30.06 in .005 22:11
BLANK TEST .000 22:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188

AUGUST 21, 2010
TIME 22:00

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.04 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklano
pqrstuvwxy{!}~@