VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 0 4 2009

DataMaster cdm S/N 130188

Supervisor/Operator Performing the Verification Procedure:

A

Name HUTT, William T. ID# 3359 Date 09/02/09

Agency AST / Anchor Point Post (ANPE) Phone # 235-8239 235-3000

Instrument Location 34115 Sterling Hwy. / P.O. Box 807

B

Alco S/N X301536 Target Value .075 High Pressure < 100 psi

Alco Test Values .077 (1st Alco) .076 (2nd Alco)

Signature William T. Hutt

9/08/09

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of Oct, 2009.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Breath Test Record

Serial Number: 130188 | Model: DataMaster | Location: APTO | AST/Anchor Point

Subject Information

Last Name: Verification Of Cal | D.L. #: 01234567

First Name: Etoh

Operator Information

Operator Name: Hutt, William T | Case: 09-00000

Operator #: 3359 | Department: AST0 / ANPE

Test Results

Alco Target	.075	Alco Serial	X301536		
Dry Gas 1 Nominal	0.075	Barometer	29.93	Test Type	V
Dry Gas 1	0.077	Time:	23:16:23	Test Date	09/02/2009
Subject Alcohol	0.000	Time:	23:18:16	Test Time	23:14:44
Dry Gas 2	0.076	Time:	23:19:39	Report	Y

Diagnostic Test Record

Serial Number: 130188 | Model: DataMaster | Location: APTO | AST/Anchor Point

Diagnostic Test Results

Test Date	09/02/2009	Computer	Okay
Test Time	23:07:41	Program	Okay
Software Date	02/20/2001		
Sample Chamber (C)	49		
Breath Tube (C)	42		
Barometer (in)	29.93		
Flow Detector	Okay		
High Speed Pump	Okay		
Detector	Okay		
Filter	Okay		
Quartz	Okay		
Calibration	Okay		

NONDRINKING SUBJECT TEST **DIAGNOSTIC CHECK**

maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a competent and qualified person conducting the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this _____ day of _____, 2009.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(CONTINUED FROM FRONT PAGE)

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

DataMaster cdm S/N 130188