

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 17 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130188
Name		HUTT, William T.	ID# 3359 Date 09/15/09
A	Agency	AST/Anchor Point Post (ANPE)	Phone # 235-8239 235-3000
Instrument Location		34115 Sterling Hwy/P.O. Box 027 Anchor Point, AK 99556	
B	Alco S/N	X124635 ✓	Target Value .080 ✓
			High Pressure 1150 psi
Alco Test Values		.084	.085 ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		William T. Hutt	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 8<sup>th</sup> day of Oct, 2009.

*Carolyn M. Noland*  
(Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130188

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130188 ✓

SEPTEMBER 14, 2009 ✓ WTHW

OPERATOR'S NAME: HUNT, WILLIAM T. *WTHW*  
OPERATOR'S NUMBER: 3359  
SUBJECT'S LAST NAME: VERIFICATION OF CAL.  
SUBJECT'S FIRST NAME/NI: ETON  
O.L. #: 01234567  
DEPT/AGENCY: ASTD / ANPE  
CASE/REPORT: 09-00000  
TEST TYPE: V  
ALCO TARGET VALUE: .000 ✓  
ALCO S/N: X124633 ✓

### — BREATH ANALYSIS —

.000 ADJUSTED FOR 30.16 in ✓  
ALCO TARGET .000 23:58  
BLANK TEST .000 23:59  
INTERNAL STANDARD VERIFIED 23:59  
ALCO TO 30.16 in .004 23:59  
BLANK TEST .000 00:00  
SUBJECT SAMPLE *WTHW* .000 00:01  
BLANK TEST .000 00:02  
ALCO TO 30.16 in .005 ✓ 00:03  
BLANK TEST .000 00:04

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130188 ✓

SEPTEMBER 14, 2009 ✓ WTHW  
TIME 23:52

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.16 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~@