

DEC 29 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130188

Supervisor/Operator Performing the Verification Procedure:

Name V. Eugene Fowler ID# 4386 Date 12/24/10

A Agency Alaska State Troopers Phone # 262-4453

Instrument Location Anchor Point Post

B Alco S/N X124635 Target Value .080 High Pressure 1000

Alco Test Values .083 .084
 1st Alco 2nd Alco

Signature *V. Eugene Fowler*

CRB
12/30/10

(OVER)

mailed
12/24/10
VE

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Nita J. Bolz
 Subscribed and sworn before me this 24 day of MAR, 2010

(Notary Seal Stamp)

Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130188

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188
DECEMBER 24, 2010 *COB 12/30/10*
OPERATOR'S NAME:
FOMLER/U/E
OPERATOR'S NUMBER: 4386
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: SLT5
CASE/REPORT: VOC
TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: X124635

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.64 in		
ALCO TARGET	.079	18:24
BLANK TEST	.000	18:25
INTERNAL STANDARD	VERIFIED	18:25
ALCO TV 29.64 in	.003	18:25
BLANK TEST	.000	18:26
SUBJECT SAMPLE <i>COB 12/30/10</i>	.000	18:27
BLANK TEST	.000	18:27
ALCO TV 29.62 in	.004	18:28
BLANK TEST	.000	18:29

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188
DECEMBER 24, 2010
TIME 18:19

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/28/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	43c
BAROMETER:	29.64 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!*M?&?()++,-./0123456789:;<=>?@BCDEFG
HIJKL NOPQRSTU VWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~!@