

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 06 2010

Supervisor/Operator Performing the Verification Procedure:	
Name <u>HUTT, William T.</u>	ID# <u>3359</u> Date <u>12/31/09</u>
Agency <u>AST/Anchor Point Post (ANPE)</u>	Phone # <u>235-8239</u> <u>235-3000</u>
Instrument Location <u>34115 Sterling Hwy/P.O. Box 807 Anchor Point AK</u> <u>99556</u>	
DataMaster cdm S/N <u>130188</u>	
Alco S/N <u>X124635</u> Target Value <u>.080</u> High Pressure <u>1100 psi</u>	
Alco Test Values <u>.084</u> ✓ <u>.085</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>William T. Hutt</u> ✓	
<u>CB</u> <u>1/5/10</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of January, 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130188

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188 ✓
DECEMBER 31, 2009 ✓
OPERATOR'S NAME: HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME: ✓
VERIFICATION OF CAL ✓
SUBJECT'S FIRST NAME/MI : ETOH ✓
O.L. #: 01234567 ✓
DEPT/AGENCY: AST0 / ANPE ✓
CASE/REPORT: 09-00000 ✓
TEST TYPE: V ✓
ALCO TARGET VALUE: .000 ✓
ALCO S/N: X124635 ✓
— BREATH ANALYSIS —
.000 ADJUSTED FOR 30.29 in ✓
ALCO TARGET .000 04:45
BLANK TEST .000 04:45
INTERNAL STANDARD VERIFIED 04:46
ALCO TV 30.29 in .004 04:46
BLANK TEST .000 04:47
SUBJECT SAMPLE .000 04:48
BLANK TEST .000 04:49
ALCO TV 30.29 in .005 04:49
BLANK TEST .000 04:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188
DECEMBER 31, 2009 WTHD
TIME 04:39
— DIAGNOSTIC CHECK —
COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 30.29 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY
PRINTER TEST
!*"#&'()*+,-./@123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■