

APR 06 2011

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130188

Supervisor/Operator Performing the Verification Procedure:  
 Name V. Eugene Forder ID# 4386 Date 3/30/11

A Agency Alaska State Troopers Phone # 262-4453  
 Instrument Location Anchorage Point Post

---

B Alco S/N X124635 Target Value .080 High Pressure 1000

Alco Test Values .082 .082  
 1st Alco 2<sup>nd</sup> Alco

Signature V E Forder

(OVER)

*Handwritten initials and date:* BMB 4/7/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Handwritten signature of Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 27 day of MAY, 2010. 2011



*Handwritten signature of Kyle Klopfer*  
 (Notary Seal Stamp)  
 Daniel L. Nicholas KYLE KLOPFER  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130188 *SMB 4/7/11*

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130188

MARCH 30, 2011 *SMB 4/7/11*

OPERATOR'S NAME:

FOULER/U/E

OPERATOR'S NUMBER: 4386

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: SLT5

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .080

ALCO S/N: X124635

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.39 in		
ALCO TARGET	.078	13:46
BLANK TEST	.000	13:47
INTERNAL STANDARD	VERIFIED	13:47
ALCO TV 29.39 in	.082	13:48
BLANK TEST	.000	13:49
SUBJECT SAMPLE	.000	13:49
BLANK TEST	.000	13:50
ALCO TV 29.39 in	.082	13:50
BLANK TEST	.000	13:51

*SMB 4/7/11*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130188

MARCH 30, 2011

TIME 13:44

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/28/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	29.39 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~