VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 17 2009

DataMaster cdm S/N 130189 (2/19/09 CMW)

Supervisor/Operator Performing the Verification Procedure:

A

Name Gerald Meece ID# 3220 Date 2-9-09

Agency St Paul DPS Phone # 907 5463130

Instrument Location City Hall

B

Alco S/N 56801 Target Value .081 High Pressure 350

Alco Test Values ~~.081~~ .075 (2/19/09 CMW) .081

1st Alco 2nd Alco

Signature Gerald Meece

CAB 2/23/09

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gerald B Meece, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

Swartz 2/23/09

Jeanne Swartz Date

Scientific Director Designee

State of Alaska

State Breath Alcohol Testing Program

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130189

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130189

FEBRUARY 09, 2009 CDB 2/23/09

OPERATOR'S NAME:
MEECE GERALD
OPERATOR'S NUMBER: 3220
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
L
O.L. #: 123456
DEPT/AGENCY: SHP1
CASE/REPORT: U
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: 56801

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.03 in

ALCO TARGET	.081	15:12
BLANK TEST	.000	15:13
INTERNAL STANDARD	VERIFIED	15:13
ALCO TV 30.03 in	.075	15:13
BLANK TEST	.000	15:14
SUBJECT SAMPLE	.000	15:15
BLANK TEST	.000	15:16
ALCO TV 30.03 in	.081	15:16
BLANK TEST	.000	15:17

CDB 2/23/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130189

FEBRUARY 09, 2009
TIME 14:57

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.03 in
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~