

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130189</u> ✓
Name	<u>Gerald Meece</u>	ID#	<u>3220</u> Date <u>5-4-09</u> ✓
Agency	<u>ST PAUL public Safety</u>	Phone #	<u>907 546 3130</u>
Instrument Location	<u>ST PAUL City Hall</u> ✓		
Alco S/N	<u>56801</u> ✓	Target Value	<u>.1081</u> ✓ High Pressure <u>380</u>
Alco Test Values	<u>.1082</u> <u>1st Alco</u>	<u>.080</u> <u>2nd Alco</u>	
Signature	<u>Gerald Meece</u> ✓		<u>CM</u> <u>5/13/09</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

Carolyn M. Noland
(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130189

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130189

MAY 04, 2009

OPERATOR'S NAME: *GB 5/12/09*
MEEDE GERALD
OPERATOR'S NUMBER: 3220
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
L
O.L. #: 103456
DEPT/AGENCY: SDF1
SUBJECT REPORT: U
TEST TYPE: U
ALCO TARGET VALUE: .001 ✓
ALCO S/N: 56001

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 30.19 in ✓
ALCO TARGET .001 12:29
BLANK TEST .000 12:30
INTERNAL STANDARD VERIFIED 12:30
ALCO TV 30.19 in .002 ✓ 12:31
BLANK TEST .000 12:32
SUBJECT SAMPLE *GB 5/12/09* .000 12:32
BLANK TEST .000 12:33
ALCO TV 30.19 in .000 ✓ 12:33
BLANK TEST .000 12:34

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130189

MAY 04, 2009 ✓
TIME 12:25

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.19 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■