

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130189</u>	
Name	<u>Gerald Meece</u>	ID#	<u>3220</u> Date <u>5-27-08</u> <sup>7<sup>6M</sup></sup> ✓
A Agency	<u>St. Paul DPS</u> ✓	Phone #	<u>907 546 3130</u>
Instrument Location <u>City Hall Bldg</u>			
Alco S/N <u>56801</u> ✓		Target Value	<u>081</u> ✓ High Pressure <u>550 psi</u>
Alco Test Values		<u>080</u> <sup>GM</sup> <u>080</u> ✓ 5/28/08 "080" ✓	
Signature <u>Gerald Meece</u>		CSB 5/14/08	

(OVER)

Under the Alaska Rules of Evidence, I certify that (if not, write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gerald B Meece, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

5/29/08

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

Date

MAY 12 2008  
B 103/00

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130189

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130189

MAY 07, 2008

OPERATOR'S NAME:

MEECE BERHALD

OPERATOR'S NUMBER: 3228

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

D.L. #: 122345

DEPT/AGENCY: SNP1

CASE/REPORT: 0012

TEST TYPE: U

ALCO TARGET VALUE: 001

ALCO S/N: 56901

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.64 in  
ALCO TARGET .000 18:36  
BLANK TEST .000 18:37  
INTERNAL STANDARD VERIFIED 18:37  
ALCO TO 29.65 in .000 18:38  
BLANK TEST .000 18:39  
SUBJECT SAMPLE .000 18:40  
BLANK TEST .000 18:41  
ALCO TO 29.65 in .000 18:41  
BLANK TEST .000 18:42

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130189

MAY 07, 2008

TIME 18:45

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/26/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.65 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxy{|}~"■