

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAY 10 2010

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130189</u>	
Name	<u>Richard Eaton</u>	ID#	<u>7287</u> Date <u>5/15/10</u>
A Agency	<u>St. Paul Island, Ak</u>	Phone #	<u>907-546-3130</u>
Instrument Location	<u>City Hall</u>		
B Alco S/N	<u>56801</u>	Target Value	<u>.081</u> High Pressure <u>200</u>
Alco Test Values	<u>.080</u>	<u>.080</u>	
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature	<u>Richard B Eaton</u>		
(OVER)			

*BMB*  
*5/24/10*

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 17<sup>th</sup> day of June, 2010.  
Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



