

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130189
Name	Gerald Meece	ID#	3220
		Date	08/01/08
Agency	St. Paul DPS	Phone #	907 546 3130
Instrument Location	City Hall Bldg ST PAUL AK ✓		
Alco S/N	56801 ✓	Target Value	.081 ✓
		High Pressure	380
Alco Test Values	.081 ✓	.081 ✓	
	1 st Alco	2 nd Alco	
Signature	Gerald Meece		08/01/08
			003 8/12/08

(OVER)


Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gerald B Meece, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

8/13/08

Date

BT9 03/00

AUG 11 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130189

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130189

AUGUST 01, 2008

OPERATOR'S NAME:
NEECE GERALD

OPERATOR'S NUMBER: 3220

SUBJECT'S LAST NAME:
UOC

SUBJECT'S FIRST NAME/MI :
L

O.L. #: 1233456

DEPT/AGENCY: SNP1

CASE/REPORT: 00000

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: 56001

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.17 in
ALCO TARGET .081 11:50
BLANK TEST .000 11:51
INTERNAL STANDARD VERIFIED 11:51
ALCO TV 30.17 in .081 11:51
BLANK TEST .000 11:52
SUBJECT SAMPLE .081 11:53
BLANK TEST .000 11:53
ALCO TV 30.19 in .081 11:54
BLANK TEST .000 11:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130189

AUGUST 01, 2008
TIME 11:42

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 41c
BAROMETER: 30.17 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~@