

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

SEP 21 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130189

Name Gerald Meece ID# 3220 Date 9-17-09

A Agency ST. PAUL DPS Phone # 907 546 3130

Instrument Location City Hall Bldg

B Alco S/N 56801 Target Value .081 High Pressure 250 *

Alco Test Values .080 .081 .080 * Also Bottle Request enclosed

Signature Gerald Meece 9/17/09

(OVER)

BMB
10/13/09


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of Oct, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130189 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130189

SEPTEMBER 17, 2009

OPERATOR'S NAME:
NEECE GERALD
OPERATOR'S NUMBER: 3226
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME: []

D.L. # 123456
SEPT/NEECE GERALD
TYPE/NEECE
TEST # 12345
SUBJECT NAME []
ALCO S/N 08801

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.70 in ✓
ALCO TARGET .000 11:44
BLANK TEST .000 11:45
INTERNAL STANDARD VERIFIED 11:45
ALCO TO 29.70 in .000 11:45
BLANK TEST .000 11:46
SUBJECT SAMPLE .000 11:47
BLANK TEST .000 11:47
ALCO TO 29.70 in .000 11:48
BLANK TEST .000 11:49

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130189

SEPTEMBER 17, 2009 ✓
TIME 11:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.70 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL/MNOPQRSTUWXYZ-12345678901234567890
abcdefghijklmnopqrstuvwxyz{|}~