

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 3 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130190 ✓

Name Bill J Ferris ID# 5004 Date 3-28-10

A Agency Togiak Police Department Phone # 493-5212 3/31/10 8/31/10

Instrument Location Togiak PD 100 main Street

B Alco S/N 57023 Target Value 081 High Pressure 900 psi

Alco Test Value Average 081 080
1st Alco 3/31/10 2nd Alco 3/31/10

Signature Bill J. Ferris ✓

(OVER) BMB
4/1/10

(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130190 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

MARCH 28, 2010

OPERATOR'S NAME: **BMB 4/11/10**
FERRIS BILL J
OPERATOR'S NUMBER: 5004
SUBJECT'S LAST NAME:
UDC
SUBJECT'S FIRST NAME/MI :
XXXXXXXX
O.L. #: XXXXXXXX
DEPT/AGENCY: T061
CASE/REPORT: XX-1234
TEST TYPE: U
ALCO TARGET VALUE: 001 ✓
ALCO S/N: 57023 ✓

— BREATH ANALYSIS —

.001 ADJUSTED FOR 29.00 in		
ALCO TARGET	.078	11:11
BLANK TEST	.000	11:12
INTERNAL STANDARD	VERIFIED	11:12
ALCO TU 29.00 in	.001 ✓	11:13
BLANK TEST	.000	11:13
SUBJECT SAMPLE	.000	11:14
BLANK TEST	.000	11:15
ALCO TU 29.00 in	.000 ✓	11:15
BLANK TEST	.000	11:16

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

MARCH 28, 2010 ✓
TIME 11:08

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	29.00 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF G
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefg h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢